



- \_\_\_\_\_ Failure to perform and/or document required daily area monitoring surveys
- \_\_\_\_\_ Failure to keep personnel informed of radionuclide use, storage and/or disposal procedures.
- \_\_\_\_\_ Survey meter or Geiger counter not present, uncalibrated or inoperative when required.
- \_\_\_\_\_ Inadequate or improper shielding provided for personnel protection.
- \_\_\_\_\_ Personnel dosimeters not in use when required.
- \_\_\_\_\_ Food and/or beverages stored, prepared or consumed in lab.
- \_\_\_\_\_ Pipetting of radioactive materials by mouth.
- \_\_\_\_\_ Unauthorized relocation of a radionuclide laboratory.

**III. Class C Violations:**

- \_\_\_\_\_ Radionuclide work areas inadequately or improperly posted.
- \_\_\_\_\_ Inadequate documentation of radionuclide use and disposal
- \_\_\_\_\_ Failure to use proper personnel protective equipment as required.

**IV. Other Violations:**

---

---

---

**V. Further description of above indicated items:**

---

---

---

---

---